

April 24th, 2019, Genua: Ferry Box Workshop 2019

Dagmar Daehne & Anne Ritzmann

Experiences with a FerryBox on a research vessel operating in the Wadden Sea




Content

- Special conditions for our FerryBox operating in the Wadden Sea
- Modifications and improvements in our FerryBox
- Sensitivity study and validation for nutrient analysers

Sensors

- Temperature (SBE 38, Seabird)
- Thermosalinograph (SBE45, Seabird)
- pH electrode (Xylem)
- Turbidity Sensor (Hach Lange)
- Oxygen Optode (Aandera)
- Algae Online Analyser (bbe)
- Wet-chemistry analysers for anorganic nutrients (Systea)




Differences to offshore operations

- FerryBox is installed on a governmental research vessel “Burchana” instead on a ferry
- Frequent changes in active and non-active times
- When the crew is off-duty, the FerryBox has to be standby and seacocks have to be closed

Operating in turbid coastal waters and estuarine rivers (Ems, Weser)




Vessel is not sailing on fixed routes like ferries: Area of operation in 2018 and 2019



FerryBox operating in the Wadden Sea

Most problems we had were created by the turbid water containing a high level of suspended matter !!

Modifications and improvements - Bypass




The bypass is used in the first 30 seconds of each track to prevent the inlet of sediments into the FerryBox system.


Scratched Cuvette in the Algae Online Analyzer



Modification of cleaning interval - Algae Online Analyzer



Installation of humidity check - Algae Online Analyzer




The screenshot displays the NLWKN control interface with the following elements:

- Control Panel:** Includes tabs for Control, Log and States, Module, Sample times, Sample positions, Position control, Wash times, Process parameter, Navigation, and Supervisor actions.
- Path of errorlog:**
- path of prozesslog:**
- washtimes.xml:**
- positions.xml:**
- time out control.xml:**
- FerryBox name:**
- Process status:**
- Command:**
- Check housekeeping data?**
- Error counter:**
- Reset error counter?**
- Numbers of errors during operate:** If the error counter=3 the system will stay in wash or standby status. Please check intake and pump.
- Track?**
- Meas delay:**
- Check AOA humidity:** **Check and go in empty:**
- Humidity in AOA:**
- Code:**
- Samplepositions.xml:**
- off shore 2:**
- harbor 2:**

Shallow Inlet for the FerryBox



Effects of missing sea chest




Effects of missing sea chest



Temp_SBE45

Effects of missing sea chest




Improvements by the sea chest




- calming the incoming water, less bubbles in the debubbler and in the pipes
- The coarse-meshed grit in front of the sea chest prevents fishes coming in
- Less data losses caused by bad housekeeping values


Improvements by the sea chest – better quality of data



Webcam for remote monitoring



COM-Port splitter (VSPE) for on site position control using QGIS



- Serial port data can be used by one application only
- With the port splitter we can use GPS-data also for Quantum GIS
- Display the exact position on the FerryBox monitor

Validation and sensitivity study for anorganic nutrients



- Analysers are situated in our laboratory
- Samples are taken by the water sampler

Sensitivity study of anorganic nutrients


Parameter	Calibration Range [$\mu\text{mol/L}$]	Limit of Linearity SYSTEBA [$\mu\text{mol/L}$]	Limit of Linearity Lab tests [$\mu\text{mol/L}$]
P-PO ₄	0 - 6,45	19,3	24
N-NH ₄	0 - 30	47,1	40
*N-NO ₃ LOW	1,12 - 35	46	50
*N-NO ₃ HIGH	20,71 - 714,3	714,3	712
N-NO ₂	1 - 7,14	9,3	11
Si-SiO ₂	0 - 17,86	24,9	30

Our limit is set within a deviation of 10 %


Sensitivity study of anorganic nutrients – measured in dilution

Parameter	Calibration Point Dilution [$\mu\text{mol/L}$]	Limit of Linearity Dilution [$\mu\text{mol/L}$]	Limit of Linearity Lab tests [$\mu\text{mol/L}$]	Deviation [%]
P- PO_4	no dil.	-	-	-
N- NH_4	30 - 76,8	142,9	120	< 10
*N- NO_3 LOW	35 - 129,5	214	-	100: 17 120: 28
*N- NO_3 HIGH	no dil.	no dil.	-	
N- NO_2	7,14 - 33	33	-	10: 40 20: 40
Si- SiO_2	17,86 - 141,6	141,6	120	<10

Nutrients: Scattering range of calibration values



Validation of nutrient analyser data at measurement campagne



- Analyses for PO₄, NH₄ and NO₂ match those from the laboratory
- NO₃: lower concentrations than lab analyses, but mostly same trend
- Analysis for SiO₂ match only in the lower range up to 120 $\mu\text{mol/l}$

Conclusions for nutrient analysis quality check

- Exact measurements for PO_4 and NH_4 (and NO_2)
- For NO_3 measurements it's good to know the approximate quantity
- NO_3 low method up to $50 \mu\text{mol/l}$ useful
- No accuracy for SiO_2 measurements in freshwater
- The trend of the concentrations are well represented
- Measuring in dilution usually results in measurement uncertainty